

BRIHANMUMBAI MUNICIPAL CORPORATON
SCHEDULE NO. IV OF 2023-25 (X-RAY FILMS, CHEMICALS, E. C. G. ROLLS, MATERIELS ETC.)
CONTRACT PERIOD :- From Dt. of Sanction Of SCR to Two Years
SCHEDULE COPY
Bid No. 7200043471

ITEM NO.	DESCRIPTION	Unit Per	Quantity & Supply PACKING	Sample Required	PRE-QUALIFYING CRITERIA
1	2	3	4	5	6
1	Polyster base round corner X-Ray Films (8" X 10") Green sensitive, blue base with Base Fog less than or equal to 0.23 & Dmax not less than 3.10 Interleaved. The density values should be met at developer Temperature 32 ^o c To 35 ^o c	Pkt. (1x 50 Films)	872 Pkt. 1x 50 Films	2 Pkt. 1x 50 Films	ISO 9001 : 2000 Certificate of Manufacturer & Slitter.
2	Polyster base round corner X-Ray Films (10" X 12") Green sensitive, blue base with Base Fog less than or equal to 0.23 & Dmax not less than 3.10 Interleaved. The density values should be met at developer Temperature 32 ^o c To 35 ^o c	Pkt. (1x 50 Films)	1181 Pkt. 1x 50 Films	2 Pkt. 1x 50 Films	ISO 9001 : 2000 Certificate of Manufacturer & Slitter.
3	Polyster base round corner X-Ray Films (12" X 15") Green sensitive, blue base with Base Fog less than or equal to 0.23 & Dmax not less than 3.10 Interleaved. The density values should be met at developer Temperature 32 ^o c To 35 ^o c	Pkt. (1x 50 Films)	2496 Pkt. 1x 50 Films	2 Pkt. 1x 50 Films	ISO 9001 : 2000 Certificate of Manufacturer & Slitter.
4	Dental X-ray Films Intraoral Periapical The density values should be met at developer blue base with Base Fog less than or equal to 0.20 & Dmax not less than 3.10 The density values should be met at developer Temperature 32 ^o c To 35 ^o c	Pkt. (1x 150 Films)	2295 Pkt. 1x 150 Films	2 Pkt. 1x 150 Films	ISO 9001 : 2000 Certificate of Manufacturer & Slitter.
5	Liquid developer Replenisher for automatic processor (starter solution, if recommended by the manufacturer, should also be supplied) (Price of item no. 8 and item no. 9 will be considered as a unit)	Cans (To make 19 to 22.5 Lit.)	326 Cans (To make 19 to 22.5 Lit.)	2 Cans (To make 19 to 22.5 Lit.)	ISO 9001 : 2000 Certificate of Manufacturer.
6	Liquid fixer hardner for automatic processor (Price of item no. 8 and item no. 9 will be considered as a unit)	Cans (To make 19 to 22.5 Lit.)	260 Cans (To make 19 to 22.5 Lit.)	2 Cans (To make 19 to 22.5 Lit.)	ISO 9001 : 2000 Certificate of Manufacturer.
7	Pigtail drainage catheters 8F 30 – 30.00 to 30.90 Cm	No.	4000 Nos.	2 Nos.	1. Drug Mfg. Lic./ Import Lic. 2. CE Cft. from notified body 3. US FDA Approved.
8	Pigtail drainage catheters 10F 30 – 30.00 to 30.90 Cm	No.	4300 Nos.	2 Nos.	1. Drug Mfg. Lic./ Import Lic. 2. CE Cft. from notified body 3. US FDA Approved.
9	INJECTION Inj. Meglumine Diatrizoate Sodium Diatrizoate Iodine content 365-375 mg/ml in 20 ml. Ampoule/Vial	Amp/ Vial (1 x 20 ml)	8100 Amp/ Vial (1 x 20 ml)	2 Amp/ Vial (1 x 20 ml)	1. Drug Mfg. Lic./ Import Lic. 2. Itemwise WHO-GMP Cft. 3. FDA Lab Test Report/ 4. EMA Cft.

ITEM NO.	DESCRIPTION	Unit Per	Quantity & Supply PACKING	Sample Required	PRE-QUALIFYING CRITERIA
1	2	3	4	5	6
10	Pigtail Angiography catheter with multiple side holes length 65 to 75 cm 4F	No.	1570 Nos.	2 Nos.	1. Drug Mfg. Lic./ Import Lic. 2. CE Cft. from notified body 3. US FDA Approved.
11	Pigtail Angiography catheter with multiple side holes length 65 to 75 cm 5F	No.	475 Nos.	2 Nos.	1. Drug Mfg. Lic./ Import Lic. 2. CE Cft. from notified body 3. US FDA Approved.
12	Inj. Gadolinium based Contrast for MRI for Paediatric use	Vial 1 x 10 ml	8300 Vial 1 x 10 ml	2 Vial 1 x 10 ml	1. Drug Mfg. Lic./ Import Lic. 2. Itemwise WHO-GMP Cft. 3. FDA Lab Test Report 4. EMA Cft.
13	Inj. Gadolinium based Contrast for MRI for Liver/Liver specific (should have Hepatobiliary excretion)	Vial 1 x 10 ml	8510 Vial 1 x 10 ml	2 Vial 1 x 10 ml	1. Drug Mfg. Lic./ Import Lic. 2. Itemwise WHO-GMP Cft. 3. FDA Lab Test Report 4. EMA Cft.
14	Inj. Gadolinium based Contrast for MRI for general use	Vial 1 x 10 ml	24900 Vial 1 x 10 ml	2 Vial 1 x 10 ml	1. Drug Mfg. Lic./ Import Lic. 2. Itemwise WHO-GMP Cft 3. Lab Test Report 4. EMA Cft.
15	Inj. Gadolinium based Contrast for MRI for Angiography	Vial 1 x 20 ml.	2200 Vial 1 x 20 ml.	2 Vial 1 x 20 ml.	1. Drug Mfg. Lic./ Import Lic. 2. Itemwise WHO-GMP Cft 3. Lab Test Report 4. EMA Cft.
16	Disp Lure Lok Syringes 3 ml. For glue injector	No.	273320 Nos.	2 Nos.	1. Drug Mfg. Lic./ Import Lic.
17	Disp Lure Lok Syringes 1 ml. For glue injector	No.	164100 Nos.	2 Nos.	1. Drug Mfg. Lic./ Import Lic.
18	X-Ray Folders Laminated 250 gsm 14 1/2" x 17 1/2"	No.	288000 Nos.	2 Nos.	----
19	Agfa Dry Star Films (DT 2B) Size 14" x 17"	Pkt. (1 x 100 Films)	14100 Pkts. (1 x 100 Films)	2 Pkts. (1 x 100 Films)	
20	Agfa Dry Star Films (DT 2B) Size 11" x 14"	Pkt. (1 x 100 Films)	5700 Pkts. (1 x 100 Films)	2 Pkts. (1 x 100 Films)	
21	Agfa Dry Star Films (DT 2B) Size 8" x 10"	Pkt. (1 x 100 Films)	10830 Pkts. (1 x 100 Films)	2 Pkts. (1 x 100 Films)	
22	Dual pack 200 ml syringe with aspirator, "Y" Shaped coiled Tube Y-tube sterile, Non-pyrogenic (with max 400 psi) compatible with dual head pressure injector (LF/Medrad/ Sea Crown)	No.	10750 Nos.	2 Nos.	1. Drug Mfg. Lic./ Import Lic.
23	Extension "Y" Shaped coiled tubing for Injector 150 cms. (-5 or +15cm) Low Pressure (Less than 400 PSI)	No.	33000 Nos.	2 Nos.	1. Drug Mfg. Lic./ Import Lic.
24	Dual Disposable Syringes (60 to 65 ml each) with "Y" Shaped coiled Tube connecting tubes small spike for contrast and large spike for saline compatible with MRI injector (LF or Madrad).	No.	2100 Nos.	2 Nos.	1. Drug Mfg. Lic./ Import Lic.

ITEM NO.	DESCRIPTION	Unit Per	Quantity & Supply PACKING	Sample Required	PRE-QUALIFYING CRITERIA
1	2	3	4	5	6
25	One way stopcock (High Pressure)	No.	6900 Nos.	2 Nos.	1. Drug Mfg. Lic./ Import Lic.
26	Low Pressure extension "Y" Shaped coiled tubing min. 150cms (-5 or +15 cms) with coiled connecting tubes & small spike for contrast & large spike for saline for Medical MRI Injector.	No.	4200 Nos.	2 Nos.	1. Drug Mfg. Lic./ Import Lic.
27	Fuji Films for C. T. Scan 14" x 17" Dry.	Pkt. (1 x 100 Films)	145 Pkts. (1 x 100 Films)	2 Pkts. (1 x 100 Films)	
28	Agfa Dry Star Films Size 10" x 12"	Pkt. (1 x 100 Films)	6002 Pkts. (1 x 100 Films)	2 Pkts. (1 x 100 Films)	
29	Konica Dry film for Dry Camera Size 8" x 10"	Pkt. (Pkt. Of 125 films)	4145 Pkts. (Pkt. Of 125 films)	2 Pkts. (Pkt. Of 125 films)	
30	Konica Dry film for Dry Camera Size 11" x 14"	Pkt. (Pkt. Of 125 films)	2681 Pkts. (Pkt. Of 125 films)	2 Pkts. (Pkt. Of 125 films)	
31	Konica Dry film for Dry Camera Size 14" x 17"	Pkt. (Pkt. Of 125 films)	2777 Pkts. (Pkt. Of 125 films)	2 Pkts. (Pkt. Of 125 films)	
32	Flavoured, ionic, water soluble, oral Contrast Iodine content 240-250 mg/ ml. In 30 ml	Vial	10300 Vial	2 Vial	1. Drug Mfg. Lic./ Import Lic. 2. Itemwise WHO-GMP Cft. 3. FDA Lab Test Report 4. EMA Cft.
33	Bernstein catheter 4 F 100 cm	No.	1400 Nos.	2 Nos.	1. Drug Mfg. Lic./ Import Lic. 2. CE Cft. from notified body 3. US FDA Approved.
34	Bernstein catheter 5 F 100 cm	No.	750 Nos.	2 Nos.	1. Drug Mfg. Lic./ Import Lic. 2. CE Cft. from notified body 3. US FDA Approved.
35	Guiding Catheter 5 F 100 cm	No.	3800 Nos.	2 Nos.	1. Drug Mfg. Lic./ Import Lic. 2. CE Cft. from notified body 3. US FDA Approved.
36	Chiba needle 21 G 15 cms length	No.	6610 Nos.	2 Nos.	1. Drug Mfg. Lic./ Import Lic.
37	Chiba needle 22 G 15 cms length	No.	7010 Nos.	2 Nos.	1. Drug Mfg. Lic./ Import Lic.
38	PTBD drainage catheter 6 F min. 45 cms, with side holes up to 10 cms from the tip.	No.	1700 Nos.	2 Nos.	1. Drug Mfg. Lic./ Import Lic. 2. CE Cft. from notified body 3. US FDA Approved.
39	PTBD drainage catheter 7 F min. 45 cms, with side holes up to 10 cms from the tip.	No.	2500 Nos.	2 Nos.	1. Drug Mfg. Lic./ Import Lic. 2. CE Cft. from notified body 3. US FDA Approved.
40	PTBD drainage catheter 8 F min. 45 cms, with side holes up to 10 cms from the tip.	No.	3000 Nos.	2 Nos.	1. Drug Mfg. Lic./ Import Lic. 2. CE Cft. from notified body 3. US FDA Approved.
41	Disposable c t scan Pressure injector syringes (190 ml syringe + quick fill tube) for Bayer Medrad salient injector (Toshiba)	No.	7500 Nos.	2 Nos.	1. USFDA Approved 2. CE Cft. from notified body
42	Extension "Y" Shaped Coild tube adapter dual check valves	No.	2600 Nos.	2 Nos.	1. Drug Mfg. Lic./ Import Lic.

ITEM NO.	DESCRIPTION	Unit Per	Quantity & Supply PACKING	Sample Required	PRE-QUALIFYING CRITERIA
1	2	3	4	5	6
43	Pigtail drainage catheter with trocar 12 F, length min. 25 cms	No.	2375 Nos.	2 Nos.	1. Drug Mfg. Lic./ Import Lic. 2. CE Cft. from notified body 3. US FDA Approved.
44	Pigtail drainage catheter with trocar 10 F, length minimum 25 cms	No.	2670 Nos.	2 Nos.	1. USFDA Approved 2. CE cft. from notified body
45	Nonionic iodixanol Isoosmolar, 320 mg 50 ml	Bottle (1 X 50 ml)	7250 Bottles (1 X 50 ml)	2 Bottles (1 X 50 ml)	1. Drug Mfg. Lic./ Import Lic. 2. Itemwise WHO-GMP Cft. 3. FDA Lab Test Report 4. EMA Cft.
46	Injection Gadolinium based Contrast 1.0 mmol for MRI for General Use	Vial (1 x 10 ml)	14500 Vial (1 x 10 ml)	2 Vial (1 x 10 ml)	1. Drug Mfg. Lic./ Import Lic. 2. WHO-GMP/COPP 3. USFDA Approved Test Report / Mfg.Test Report
47	Injection Gadolinium based Contrast 1.0 mmol for MRI for Angiography	Vial (1 x 7.5ml)	1800 Vial (1 x 7.5ml)	2 Vial (1 x 7.5ml)	1. Drug Mfg. Lic./ Import Lic. 3. WHO-GMP/COPP 3. USFDA Approved Test Report / Mfg.Test Report
48	Disposable MRI syringes kit for 65/115 MR Injector system 65 ml contrast syringes and 115 ml NS syringes with connecting tubing Combo pack for Spectris solaris Medrad Pressure injector	No.	2000 Nos.	2 Nos.	1. USFDA Approved 2. CE cft. from notified body
49	Spring loaded fully automatic and semi automatic biopsy instrument with adjustable penetration depth of 18 mm And 25 mm. Having trocar tip with dimpling on needle and fire ready indicator. 14 g & 10 cm in length	No.	720 Nos.	2 Nos.	1. USFDA Approved 2. CE cft. from notified body
50	Spring loaded fully automatic and semi automatic biopsy instrument with adjustable penetration depth of 18 mm And 25 mm. Having trocar tip with dimpling on needle and fire ready indicator. 16 g & 10 cm in length	No.	45 Nos.	2 Nos.	1. USFDA Approved 2. CE cft. from notified body
51	Coil shape breast tissue marker with both front trigger & an alternate rear trigger, should be visible in Ultrasound, Mammography and MRI	No.	400 Nos.	2 Nos.	1. USFDA Approved 2. CE cft. from notified body
52	Ribbon shape breast tissue marker with both front trigger & an alternate rear trigger, should be visible in Ultrasound, Mammography and MRI	No.	400 Nos.	2 Nos.	1. USFDA Approved 2. CE cft. from notified body
53	Micropuncture transitionless access set: 4 Fr-21Gauge, 7 cm length , ecogenic Tip, Nitinol and stainless steel wire with platinum tip and palladium tip	No.	575 Nos.	2 Nos.	1. Drug Mfgr Lic / Import Lic. 2. USFDA Approved
54	Micropuncture transitionless access set: 5 Fr-21Gauge, 7 cm length , ecogenic Tip, Nitinol and stainless steel wire with platinum tip and palladium tip	No.	675 Nos.	2 Nos.	1. Drug Mfgr Lic / Import Lic. 2. USFDA Approved
55	Micropuncture transitionless access set: 4 Fr-21Gauge, 4 cm length , ecogenic Tip, Nitinol and stainless steel wire with platinum tip and palladium tip	No.	535 Nos.	2 Nos.	1. Drug Mfgr Lic / Import Lic. 2. USFDA Approved
56	Micropuncture transitionless access set: 5 Fr-21Gauge, 4 cm length , ecogenic Tip, Nitinol and stainless steel wire with platinum tip and palladium tip	No.	415 Nos.	2 Nos.	1. Import Lic. 2. USFDA Approved
57	Extension coiled tube with integral T Connector 150 cm long for spectris solaris injector (low pressure) and Integrated check valve	No.	3000 Nos.	2 Nos.	1. USFDA Approved 2. CE cft. from notified body
58	Disposable pressure injector 150 ml fast turn syringe with quick fill tube for bayer Medrad Mark 7 Arterion injector	No.	2000 Nos.	2 Nos.	1. USFDA Approved 2. CE cft. from notified body
59	Extension coiled tube , 150 cm, Low pressure 300 PSI, sterile pyrogen free for Bayer Medrad salient single head injector	No.	3000 Nos.	2 Nos.	1. USFDA Approved 2. CE cft. from notified body

ITEM NO.	DESCRIPTION	Unit Per	Quantity & Supply PACKING	Sample Required	PRE-QUALIFYING CRITERIA
1	2	3	4	5	6
60	Extension Y-Tube with Dual Check Valve 300 PSI 150 cm Long Sterile, Pyrogen free For Bayer Medrad Salient Injector (Toshiba)	No.	26000 Nos.	2 Nos.	1. USFDA Approved 2. CE cft. from notified body
61	Introducer system-co-axial 6f sheath with RO Marker & 4F, 20 cms dilator with Locking, stiffening cannula. 1 no. Of 21G, 15cms needle with stylet. 0.018" nitinol wire	No.	2400 Nos.	2 Nos.	1. USFDA Approved 2. CE cft. from notified body
62	0.018" fibred Platinum Coils for embolization (Assorted sizes)	No.	600 Nos.	2 Nos.	1. USFDA Approved 2. CE cft. from notified body
63	0.035" fibred Platinum coils for embolization (Assorted sizes)	No.	175 Nos.	2 Nos.	1. USFDA Approved 2. CE cft. from notified body
64	Pedal access set 4F, 21G, 4 cms length, echogenic tip, nitinol & stainless steel wire with Platinum Tip & Palladium Tip	No.	450 Nos.	2 Nos.	1. USFDA Approved 2. CE cft. from notified body
65	Pedal access set 5F, 21G, 4 cms length, echogenic tip, nitinol & stainless steel wire with Platinum Tip & Palladium Tip	No.	300 Nos.	2 Nos.	1. USFDA Approved 2. CE cft. from notified body
66	18 G diamond tip puncture needle	No.	700 Nos.	2 Nos.	1. USFDA Approved 2. CE cft. from notified body
67	Introducer needle 18G , 7 Cms, long type pink with recommended guide wire 0.038" or smaller	No.	13500 Nos.	2 Nos.	1. USFDA Approved 2. CE cft. from notified body
68	Verterbral Biospy needle 11G, 9 cms	No.	705 Nos.	2 Nos.	1. USFDA Approved 2. CE cft. from notified body
69	Verterbral Biospy needle 11G, 15 cms	No.	605 Nos.	2 Nos.	1. USFDA Approved 2. CE cft. from notified body
70	Verterbral Biospy needle 13G, 9 cms	No.	725 Nos.	2 Nos.	1. USFDA Approved 2. CE cft. from notified body
71	Verterbral Biospy needle 13G, 15 cms	No.	625 Nos.	2 Nos.	1. USFDA Approved 2. CE cft. from notified body
72	SIM 2-4F-0.038" sterile braided angio catheter	No.	1250 Nos.	2 Nos.	1. USFDA Approved 2. CE cft. from notified body

Note: Rate should be quoted as per column No.3 (Inclusive of All Taxes)

sd/-
Dy. Dean (CPD)